

Gendex Orthoralix 9200

Integration Instructions for ViewPoint



The Gendex Orthoralix 9200 direct digital pan/ceph system, when set up correctly, will function the same as a regular scanner within VP Imaging. No additional hardware or software is needed.


Setup

Install the Gendex Orthoralix, appropriate software, and corresponding TWAIN driver. This should be accomplished by someone from Gendex.

For information on setting up VP Imaging, consult the VP Imaging documentation and contact Ortho2 Software Support with any questions.

Using the Integration

- 1 Open a patient record in ViewPoint and click the **VP Imaging** tab.
- 2 Select the appropriate layout.
- 3 Click **Scan** to open the *VP Imaging - Scan* window.
- 4 Select the appropriate timepoint.
- 5 Select the appropriate image type.
- 6 Click **Select Scanner** to open the *Select Scanner* window.
- 7 On the **Twain** tab, select the Gendex device. Click **OK**.
- 8 Click **Scan** on the *VP Imaging - Scan* window. This will start your scanning software. For more information on proceeding, refer to the Gendex documentation.
- 9 A **Count=1** notation will appear beneath the **Close** button. Scan additional images or click **Close**.



Refer to the VP Imaging documentation for instructions on how to use the Image Editor.

Contact

Orthoralix 9200
Gendex Dental Systems
340 East Route 22
Lake Zurich, IL 60047
www.gendex.com
(888) 275-5286

ViewPoint
Ortho2
1107 Buckeye Avenue
Ames, IA 50010
www.ortho2.com
(800) 346-4504 or (515) 233-1026